



Certificate of Calibration

AS LEFT

Certificate ID: L1T10655051321



For AS Found refer to; F1T10655051121

Manufacturer	Carl Zeiss
Model	CenterMax 11/12/7 Navigator
Serial Number	135391
Customer ID	-
Customer	Miltera Machining Research Corp. 60 Struck Crt. Cambridge, Ontario N1R 8L2 CA
Job Number	T10655
Calibration Certificate page	1 of 19
Date of Calibration	2021-05-13
Calibration Procedure	ISO-10360
Uncertainty of Length Measurement = $0.1 + L/1600$ [μm]	Temperature 20.38 °C

The equipment identified in this certificate was calibrated with standards that are traceable to national metrology institutes (e.g. NIST) through calibration laboratories accredited to ISO 17025. All results are reported in units of measure as defined by the International System of Units (SI).

A2LA is a signatory to several bilateral and multilateral recognition agreements. These agreements facilitate the acceptance of test and calibration data between A2LA-accredited laboratories and 46 economies around the globe. In addition, A2LA has recognition from over 30 federal, state and local government agencies, companies and associations.

The user is responsible for definition of appropriate intervals of calibration.

Please store in a secure location. Elliott-Matsuura Canada Inc. is not responsible for loss of data

UAWT Version 9.6.0.752

Calibration certificates without signature are not valid.

Signature / Name
Matthew Joyce

Date
2021-05-13

The most noteworthy contributor to the uncertainty budget of a CMM is the deviation of temperature away from the standard of 20 degrees Celsius. This calibration certificate shall not be reproduced, except in full, without written approval of Elliott-Matsuura Canada Inc. Unless otherwise annotated in protocol results, machine condition is in good working order. The reported results relate only to the item(s) specified above. All measurements performed at 95% confidence level (K=2).

1. Calibration task

Indication error E_0 for length measurements and probing error $PFTU$ are measured on the coordinate measuring machine for sensor systems with scanning capability, scanning probing error THP and the time for the scanning test are calibrated.

For sensor systems with scanning capability, roundness form measuring error **RONt (MZCI)** was measured. If a rotary table is installed, the four-axis errors **FR**, **FT** and **FA** were measured if this measurement was ordered.

The coordinate measuring machine had the following configuration at the time of calibration:

Controller:	C99N (850Mhz) #AK003217 / FW: 24.08
Probe:	VAST Gold #0017UABS
Measurement SW:	CALYPSO 2019
Reference sphere:	#R5043 r=14.9890
Rotary table:	-
X measuring range:	1100mm
Y measuring range:	1200mm
Z measuring range:	700mm

2. Calibration procedure

Calibration of the metrological features of the coordinate measuring machine was performed according to Carl Zeiss IMT procedure CL-1001. This procedure is established and validated using international metrological methods.

Efforts were made to achieve ISO 10360-2, 6.3.2 which states "*The longest calibrated test length for each position shall be at least 66 % of the maximum travel of the CMM along a measurement line through the calibrated test length.*"

When unable, $\geq 66\%$ of each linear axis is measured, via additional measurements if necessary.

The roundness form measurement errors **RONt (MZCI)** were determined by measuring a master ring in the scanning mode with $D = 50\text{mm}$.

The four-axis measurement deviations **FR** (radial), **FT** (tangential) and **FA** (axial) were determined using two ceramic spheres with $D = 30\text{mm}$. The ceramic spheres were clamped with a horizontal distance from the rotary axis of $r = 206\text{mm}$ and a horizontal distance of $d = 412\text{mm}$ as well as a vertical distance of $h = 206\text{mm}$.

The single stylus form error P_{FTU} as well as the scanning probing error THP and their test duration were determined on a ceramic sphere with $D = 25 \text{ mm}$.

The customer is identified on page 1. The machine location is specified in section 3.

The calibration standards used are specified in the relevant sections of the measurement result documentation.

3. Ambient conditions

The maximum temperature deviation from 20°C during the measurement was 0.67°C.
The calibration was performed on-site. The coordinate measuring machine is installed at the following location:

Miltera Machining Research Corp.
60 Struck Crt.
Cambridge, Ontario N1R 8L2 CA

QA Lab

4. Measurement results

The measurement results apply only to the specified time of measurement. They also apply only to the relevant installation site and machine settings at the time of calibration. All settings and correction values were documented by the calibration laboratory.

4.1 Indication error for length measurements E_0

The following parallel and stepped gauge blocks are used to determine indication errors:

GCS number: E44 valid to 2021-07-13

The following temperature measuring device was utilized to perform temperature measurement to calculate deviation from a reference temperature of 20°C (if applicable).

GCS number: E32 valid to 2023-01-22

The determined indication errors E_0 and the maximum permissible indication error for length measurements $E_{0,MPE}$ are represented in the following diagrams.


The maximum permissible indication error amounts to:

$$MPE_{E_0} = +/- (A + L/K) \leq B \text{ (L in mm)}$$


$$MPE_{E_0} = +/- (1.40 + L/300) \mu\text{m}$$

The measured results were determined with a stylus L = 94 mm and D = 12.0 mm.


indication error in pos. 1 (X axis (bottom))




Measuring length L in mm		Deviations in mm		
nominal value	actual value	mean value	minimum	maximum
20.0045	20.0043	-0.0002	-0.0002	-0.0002
179.9338	179.9345	0.0007	0.0007	0.0007
379.9153	379.9152	-0.0002	-0.0002	-0.0002
539.8482	539.8481	-0.0001	-0.0001	0.0000
699.9049	699.9055	0.0006	0.0006	0.0007

indication error in pos. 2 (X axis (top))

Measuring length L in mm		Deviations in mm		
nominal value	actual value	mean value	minimum	maximum
20.0045	20.0042	-0.0003	-0.0004	-0.0003
179.9338	179.9338	0.0000	-0.0001	0.0000
379.9153	379.9147	-0.0006	-0.0007	-0.0006
539.8482	539.8476	-0.0006	-0.0007	-0.0006
699.9049	699.9048	-0.0001	-0.0001	-0.0001


indication error in pos. 3 (Y axis (front-right))

Measuring length L in mm		Deviations in mm		
nominal value	actual value	mean value	minimum	maximum
20.0045	20.0045	0.0000	-0.0001	0.0000
179.9338	179.9336	-0.0002	-0.0002	-0.0002
379.9153	379.9155	0.0002	0.0001	0.0002
539.8482	539.8485	0.0003	0.0002	0.0003
699.9049	699.9049	0.0000	-0.0001	0.0001

indication error in pos. 4 (Y axis (rear-left))


Measuring length L in mm		Deviations in mm		
nominal value	actual value	mean value	minimum	maximum
20.0045	20.0040	-0.0005	-0.0005	-0.0005
179.9338	179.9335	-0.0003	-0.0004	-0.0003
379.9153	379.9150	-0.0003	-0.0004	-0.0003
539.8482	539.8483	0.0001	0.0000	0.0001
699.9049	699.9050	0.0002	0.0001	0.0002

indication error in pos. 5 (Z axis)




Measuring length L in mm		Deviations in mm		
nominal value	actual value	mean value	minimum	maximum
20.0045	20.0042	-0.0003	-0.0004	-0.0003
139.9404	139.9403	-0.0001	-0.0001	0.0000
259.9603	259.9602	-0.0001	-0.0001	0.0000
379.9153	379.9152	-0.0001	-0.0002	0.0000
499.9018	499.9017	-0.0001	-0.0001	-0.0001

indication error in pos. 6 (Spatial (front-right))




Measuring length L in mm		Deviations in mm		
nominal value	actual value	mean value	minimum	maximum
20.0045	20.0044	-0.0001	-0.0002	-0.0001
179.9338	179.9340	0.0002	0.0002	0.0002
379.9153	379.9153	0.0000	-0.0001	0.0000
539.8482	539.8478	-0.0005	-0.0005	-0.0004
699.9049	699.9041	-0.0008	-0.0008	-0.0008

indication error in pos. 7 (Spatial (front-left))




Measuring length L in mm		Deviations in mm		
nominal value	actual value	mean value	minimum	maximum
20.0045	20.0048	0.0003	0.0002	0.0004
179.9338	179.9343	0.0005	0.0004	0.0005
379.9153	379.9150	-0.0004	-0.0004	-0.0003
539.8482	539.8476	-0.0006	-0.0007	-0.0005
699.9049	699.9051	0.0003	0.0002	0.0003

indication error in pos. 8 (Spatial (rear-left))



Measuring length L in mm		Deviations in mm		
nominal value	actual value	mean value	minimum	maximum
20.0045	20.0044	-0.0002	-0.0002	-0.0001
179.9338	179.9335	-0.0003	-0.0004	-0.0003
379.9153	379.9148	-0.0005	-0.0005	-0.0005
539.8482	539.8478	-0.0004	-0.0005	-0.0004
699.9049	699.9046	-0.0002	-0.0003	-0.0002

indication error in pos. 9 (Spatial (rear-right))



Measuring length L in mm		Deviations in mm		
nominal value	actual value	mean value	minimum	maximum
20.0045	20.0041	-0.0004	-0.0004	-0.0004
179.9338	179.9331	-0.0007	-0.0008	-0.0006
379.9153	379.9153	0.0000	-0.0001	0.0001
539.8482	539.8485	0.0002	0.0001	0.0003
699.9049	699.9044	-0.0005	-0.0006	-0.0004

4.2 Probing error P_{FTU}

The following ceramic sphere was used to determine the probing error:

GCS number: E57 valid to 2021-06-12

The max. permissible value : **1.4 μm**

Measured value: **$P_{FTU} = (0.4 \pm 0.09) \mu\text{m}$**

Position of test object: X = 483mm Y = -750mm Z = -408mm

Temperature of test object in °C: 19.80

The measured results were determined with a stylus L = 94 mm and a D of 12.0 mm.

4.3 Scanning probing error THP and scanning test duration

The following ceramic sphere was used to determine scanning probing error THP and the scanning test duration:

GCS number: E57 valid to 2021-06-12

The max. permissible value for scanning probing error THP: **2.4 µm**

The measured value of the : **THP = (1.7 ± 0.09) µm**

The max. permissible value for the scanning test duration is: **29 s ***

The measured value : **= (26 ± 0.90) s ***

Position of test object: X = 483mm Y = -750mm Z = -408mm

Temperature of test object in °C: 19.89

The measured results were determined with a stylus L = 50 mm and a D of 3.0 mm.

*) NOTE: This measurement is outside the scope of this laboratory's A2LA scope of accreditation

4.4 Roundness form measurement error RONt (MZCI)

The following master ring is used to determine roundness form measurement error **RONt (MZCI)**:

GCS number: E11 valid to 2023-01-22

Max. permissible roundness error: $t = 1.7 \mu\text{m}$


Measurement results:

In the X/Y plane: $t = (0.4 \pm 0.14) \mu\text{m}$ ($T = 20.6^\circ\text{C}$; stylus L = 94mm and D = 12.0mm)


In the X/Z plane: $t = (0.5 \pm 0.14) \mu\text{m}$ ($T = 20.4^\circ\text{C}$; stylus L = 94mm and D = 12.0mm)

In the Y/Z plane: $t = (0.4 \pm 0.14) \mu\text{m}$ ($T = 20.5^\circ\text{C}$; stylus L = 94mm and D = 12.0mm)


Roundness form measurement error RONt (MZCI) XY-Plane



Roundness form measurement error RONt (MZCI) YZ-Plane



Roundness form measurement error RONt (MZCI) XZ-Plane



5. Measuring uncertainty

The greatest measuring uncertainty experienced during calibration is reported on page 1 of the calibration certificate.

Zero Guard Banding: Uncertainty of measurement was not factored when determining compliance to relevant specifications.

The uncertainty of measurement represents an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of K=2. The probability that the value of the measured variable will lie within the allocated value interval is 95%.

6. Certificate of conformity

If confirmed below, the coordinate measuring machine fulfills the specified requirements.

The performance of the coordinate measuring machine has been calibrated according to the relevant specifications.

[The coordinate measuring machine meets the original manufacturer's specification.](#)

7. Opinions / Interpretations resulting from calibration

Refer to the following Appendices:

1. Appendix A for MPE_{PFTU} results
2. Appendix B for MPE_{THP} results



ZEISS CALYPSO

6.8.16

Part name	MPE-P-C44-V070523-DuraMax				
Drawing number					Last 1 measurements
Order number	Appendix A				► Approval ≠ Blocked
Variant					
Company			Part ident		n.a93
Department			Time/Date		5/13/2021 2:59 PM
CMM Type	CENTERMAX		Run		All Characteristics
CMM No.	135391		No. measured values		37
Operator	Master		No. values: red		0
Text					Measurement Duration 00:01:36.0

Name	Measured value	Nominal value	+Tol	-Tol	Deviation	+/-
Temperature Compensation	19.80					
Temperature AVG START	19.79960	0.00000			19.79960	
MPE-P Value	0.00036	0.00000	0.00140	-0.00140	0.00036	
X	483.29225	0.00000			483.29225	
Y	-750.85329	0.00000			-750.85329	
Z	-408.10216	0.00000			-408.10216	
Time	64.00000	64.00000			0.00000	
TimePerProbing	2.60000	2.60000			0.00000	
Minimaler Radius	12.49765	12.49798	0.00250	-0.00250	-0.00033	
Maximaler Radius	12.49801	12.49798	0.00250	-0.00250	0.00004	
Radius(1)	12.49784	12.49798	0.00140	-0.00140	-0.00013	
Radius(2)	12.49787	12.49798	0.00140	-0.00140	-0.00010	
Radius(3)	12.49794	12.49798	0.00140	-0.00140	-0.00004	
Radius(4)	12.49790	12.49798	0.00140	-0.00140	-0.00007	
Radius(5)	12.49775	12.49798	0.00140	-0.00140	-0.00023	
Radius(6)	12.49780	12.49798	0.00140	-0.00140	-0.00018	
Radius(7)	12.49790	12.49798	0.00140	-0.00140	-0.00007	
Radius(8)	12.49765	12.49798	0.00140	-0.00140	-0.00033	
Radius(9)	12.49773	12.49798	0.00140	-0.00140	-0.00025	
Radius(10)	12.49792	12.49798	0.00140	-0.00140	-0.00005	
Radius(11)	12.49781	12.49798	0.00140	-0.00140	-0.00017	
Radius(12)	12.49796	12.49798	0.00140	-0.00140	-0.00001	
Radius(13)	12.49788	12.49798	0.00140	-0.00140	-0.00010	
Radius(14)	12.49789	12.49798	0.00140	-0.00140	-0.00009	



ZEISS CALYPSO

6.8.16

Part name MPE-P-C44-V070523-DuraMax
Order number Appendix A
Part ident n.a93
Operator Master
Time/Date 5/13/2021 2:59 PM

Name	Measured value	Nominal value	+Tol	-Tol	Deviation	+/-
Radius(15)	12.49788	12.49798	0.00140	-0.00140	-0.00010	
Radius(16)	12.49775	12.49798	0.00140	-0.00140	-0.00022	
Radius(17)	12.49780	12.49798	0.00140	-0.00140	-0.00018	
Radius(18)	12.49789	12.49798	0.00140	-0.00140	-0.00008	
Radius(19)	12.49801	12.49798	0.00140	-0.00140	0.00004	
Radius(20)	12.49785	12.49798	0.00140	-0.00140	-0.00013	
Radius(21)	12.49782	12.49798	0.00140	-0.00140	-0.00016	
Radius(22)	12.49770	12.49798	0.00140	-0.00140	-0.00028	
Radius(23)	12.49778	12.49798	0.00140	-0.00140	-0.00020	
Radius(24)	12.49786	12.49798	0.00140	-0.00140	-0.00012	
Radius(25)	12.49787	12.49798	0.00140	-0.00140	-0.00011	
Temperature AVG END	19.81350	0.00000			19.81350	
Roundness1	0.00029	0.00000	0.00140	0.00000	0.00029	



ZEISS CALYPSO

6.8.16

Part name	MPE-THP-C46-V0804010-DuraMax				
Drawing number	VAST __VAST Speed: 25/25				
Order number	Appendix B				
Variant	Last 1 measurements ► Approval ≠ Blocked				
Company	Part ident 273				
Department	Time/Date 5/13/2021 3:02 PM				
CMM Type	Run All Characteristics				
CMM No.	135391 14				
Operator	No. measured values 14 Master No. values: red 0				
Text	Measurement Duration 00:00:57.0				

Name	Measured value	Nominal value	+Tol	-Tol	Deviation	+/-
^Temperature Compensation	19.89					
Temperature AVG START	19.89485	0.00000			19.89485	
DIN EN ISO 10360-4						
Result 1: MPE-THP	0.00166	0.00240	0.00000	-0.00240	-0.00074	
Result 2: MPTtau	26.00000	29.00000	0.00000	-29.00000	-3.00000	
DIN EN ISO 10360-4 ► Service Information						
Form error 'Sphere THP' filtered	0.00166	0.00000	0.00240	0.00000	0.00166	
Form error 'Sphere THP' unfiltered	0.00593	0.00000	1000.00000	0.00000	0.00593	
Rnom-Ract	0.00116	0.00000	2.40000		0.00116	
R_min	12.49682	12.49798	0.00240	-0.00240	-0.00116	
R_max	12.49887	12.49798	0.00240	-0.00240	0.00090	
Ø _{GG} Diameter Sphere THP	24.99491	24.99595			-0.00104	
X X-Value 'Sphere THP'	483.29197	0.00000			483.29197	
Y Y-Value 'Sphere THP'	-750.85457	0.00000			-750.85457	
Z Z-Value 'Sphere THP'	-408.10187	0.00000			-408.10187	
Temperature AVG - END	19.90680	0.00000			19.90680	